## DEPARTMENT OF ELECTRICAL ENGINEERING INDIAN INSTITUTE OF ENGINEERING SCIENCE AND TECHNOLOGY, SHIBPUR, HOWRAH-711 103.

*No.* 14(83R)/2019/EE-3/21(DJG)

From: The Head of the Department, Electrical Engineering,

IIEST, Shibpur, Howrah-711 103

To: Enlisted vendors of the institute and other interested parties/ For Website Tender.

Dear Sir(s),

Sealed quotations are invited for supply of the following item(s) within **7 days** from the date of publication of this advertisement in the website. The quotation should include the 5% GST only as per institute rule, delivery charges, entry tax if any, etc. to Department of Electrical Engineering, Indian Institute of Engineering Science and Technology, Shibpur and should mention a firm delivery period. Preferences will be given to the suppliers who can supply ex-stock.

The vendors, who are not enlisted in the Institute register, should submit the copies of their valid Trade License, GST registration, PAN, latest Income Tax / Sales Tax Statement /Return, SSI/MSME certificate, if any etc. and any other commercial credentials. The institute will provide concessional GST rate certificate with the purchase order and will pay 5% GST only.

Yours faithfully,

Dated: 24/01/2019

Please put your digital/scanned signature

Signature of the indenting Officer/ Concerned Faculty Member *Prof. & Head of EE Dept.*IIEST, Shibpur, Howrah – 711 103

## **List of Items:**

Item#	Name of the item	Qty	Specifications
ESL#2	True RMS Multi-meter	04	<ul> <li>4000 Count TRUE- RMS Auto-Ranging Digital Multi-meter with large backlit LCD Display with 1" digits</li> <li>Built-in Type K thermometer for temperature measurements</li> <li>Low current measurements down to 0.1µA</li> <li>Input fuse protection</li> <li>Mis-connection warnings</li> <li>Autoranging</li> <li>Capacitance measurement</li> <li>Frequency measurement</li> <li>Complete with CAT III test leads, multi-position tilt stand and velcro strap for hanging, protective holster with test lead holder, bead wire temperature probe and 9V battery</li> <li>3 years warranty</li> <li>Complete documentation</li> </ul>
ESL#5	Laboratory DC PSU	02	Input: 230 V +/- 10%, 50 Hz single phase AC; Output: +/- 15 V (+/-2%), 3 A and +/- 5 V (+/-2%), 3 A; Output voltage regulation: no load to full load < 2%; 15 V set and 5 V set of outputs should be isolated. Weight < 1 kg, Warranty: 1 year or more