

DEPARTMENT OF METALLURGY AND MATERIALS ENGINEERING
Indian Institute of Engineering Science and Technology, Shibpur
Howrah-711 103

Enquiry No. 07/MET/IEST/2017-18

Date: August 4, 2017

To,
All the Enlisted Suppliers/Dealers/Manufacturers

Sealed quotations are invited for supply of the following Equipment for the Dept. of Metallurgy and Materials Engineering. The sealed quotations should be addressed to the Head of the Department indicating enquiry no. with date on the top of the sealed envelope. The sealed quotations should be submitted at the office of the department on any of the working day up to 21/08/2017 till 5.00 P.M.

Specifications pH meter

pH range: 0.000 to 14.999; mV range: 1999.9;
Temperature range: minimum 0 deg to 100 degC;
pH resolution: 0.001 and pH accuracy: +/- 0.002;
mV resolution and accuracy: 0.1 mV and +/-0.2 mV;
Temperature resolution and accuracy: 0.1deg and \pm 0.2 degC; minimum three points calibration system;
communication: RS-232 for printing purposes,
Ambient Operating Temperature: 5 to 45°C;
Custom LCD display.
Warranty: 3 years



(Prof. Sumit Ghosh)
Head

Prof. Sumit Ghosh
Head
Dept. Of Met. & Mat. Engg.
Indian Institute of Engineering
Science and Technology, Shibpur
Howrah-711 103

Copy forwarded for information to:

Institute Website, IEST, Shibpur, Howrah-711 103

For details the vendors may contact Dr. Sukumar Kundu (Mobile No. 8334043609) during working hours.