



OFFICE OF THE DEAN ADMINISTRATIVE AFFAIRS
Indian Institute of Engineering Science and Technology, Shibpur
(Formerly Bengal Engineering and Science University, Shibpur)
P.O.: Botanical Garden
Howrah – 711 103

Date: 25/04/2017

Corrigendum

Tender ID : 2017_IEST_194929_1

Tender Reference Number : e-Proc/CHST_18042017/ HEALTHCARE_IEST/11
dtd.18/04/2017

Kindly note that :

In the CPP Portal , In the specification of Table top refrigerated Centrifuge rotor specification is as below:

- Fixed angle rotor of 6 x 50 ml tubes or more with adapter for 50 ml (conical). and 15 ml.(conical). Rotor should be metallic and autoclavable. Speed should be around 10000 rpm or more.
- Swing out rotor with 4 x 100ml bucket with adapter and lid for 15ml and 50ml conical tube, 4500rpm
- Built-in condensation drain to eliminate condensed water is not mandatory.

Sd/-
Dean, Administrative Affairs
IEST, Shibpur