



OFFICE OF THE DEAN ADMINISTRATIVE AFFAIRS
Indian Institute of Engineering Science and Technology, Shibpur
(Formerly Bengal Engineering and Science University, Shibpur)
P.O.: Botanical Garden
Howrah – 711 103

Date: 31/05/2017

Corrigendum

Corrigendum No. Corrigendum31052017/2017_IEST_205408_1

Tender ID : 2017_IEST_205408_1

Tender Reference Number: e-Proc/CHST_23052017/ PROBE_IEST/20
dated 23-05-2017

Kindly note that :

In the CPP Portal , the technical specification of Probe sonicator following changes are requested:

1. **Output** should be read as up to 500watts, **Frequency:** 20-30kHz
2. **Probe:** Should be ideal for sample volume 1mL to 50ml by using various replaceable probes
3. **Cuphorn** is an optional accessory (for indirect sonication of 6-8 eppendorfs). The price may be quoted without cuphorn. However the price of cuphorn should be indicated and uploaded separately. The L1 will be considered without cuphorn at present. This is requested because not all the models have cuphorn facility.

Sd/-
Dean, Administrative Affairs
IEST, Shibpur